

*Sister Rod Destructive Examinations (FY20)*

# ***Appendix D: Fission Gas, Fuel Burnup, and Fuel Isotopic Measurements***

## **Spent Fuel and Waste Disposition**

*Prepared for  
US Department of Energy  
Spent Fuel and Waste Science  
and Technology*

*Oak Ridge National Laboratory  
Bruce Bevard, Joe Giaquinto,  
Cole Hexel, Germina Ilas,  
Rose Montgomery,  
Robert N. Morris,  
Benjamin Roach*

**November 30, 2020**

**M2SF-21OR010201032**

**ORNL/SPR-2020/1770**

Approved for public release  
Distribution is unlimited

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

## SUMMARY

This report documents work performed under the Spent Fuel and Waste Disposition's Spent Fuel and Waste Science and Technology program for the US Department of Energy (DOE) Office of Nuclear Energy (NE). This work was performed to fulfill Level 2 Milestone M2SF-21OR010201032, "ORNL High Burnup Confirmatory Demo Sibling Rod Testing Results," within work package SF-21OR01020103 and is an update to the work reported in M2SF-19OR0010201026 and M2SF-19OR010201028.

As a part of DOE NE High Burnup Spent Fuel Data Project, Oak Ridge National Laboratory (ORNL) is performing destructive examinations (DEs) of high burnup (HBU) (>45 GWd/MTU) spent nuclear fuel (SNF) rods from the North Anna Nuclear Power Station operated by Dominion Energy. The SNF rods, called *sister rods* or *sibling rods*, are all HBU and include four different kinds of fuel rod cladding: standard Zircaloy-4 (Zirc-4), low-tin (LT) Zirc-4, ZIRLO, and M5. The DEs are being conducted to obtain a baseline of the HBU rod's condition before dry storage and are focused on understanding overall SNF rod strength and durability. Composite fuel and defueled cladding will be tested to derive material properties. Although the data generated can be used for multiple purposes, one primary goal for obtaining the post-irradiation examination data and the associated measured mechanical properties is to support SNF dry storage licensing and relicensing activities by (1) addressing identified knowledge gaps and (2) enhancing the technical basis for post-storage transportation, handling, and subsequent disposition.

This appendix documents the status of the ORNL Phase 1 DE activities related to fission gas sample analysis, fuel burnup analysis, and fuel isotopic analysis of selected sister specimens in Phase 1 of the sister rod test program.

Table DS-1 provides a summary of the DE status.

**Table DS-1. DE Status.**

Planned DE		Status	Comments
DE.01	Collect fission gas samples and analyze	Complete	Fission gas samples were collected and analyzed. Results are consistent with publicly available database. Code-predicted fission gas production is not available; therefore, the fission gas release ratio is not available. ORNL and Pacific Northwest National Laboratory (PNNL) fission gas analyses are consistent with one another, and the data are as expected when differences in fission gas partial pressure are considered.
DE.03	Analyze fuel burnup to confirm predicted and extrapolated values	In progress	Six specimens were sent to the ORNL Radiochemical Engineering Development Center for burnup analysis (Nd, U, Pu only). Three are complete. Additionally, other sponsors are funding isotopic analyses of additional sister rod specimens (~51 isotopes measured).

This page is intentionally left blank.

## ACKNOWLEDGMENTS

Many thanks to our US Department of Energy Office of Nuclear Energy sponsor, Ned Larson, along with the Spent Fuel and Waste Science and Technology storage and transportation program leadership for their continued support. The sister rod project would not have been possible without the vision and support of the Electric Power Research Institute, Westinghouse, Framatome, and Dominion Energy.

This work would not have been possible without the support and expertise provided by the leadership and staff members of the Oak Ridge National Laboratory's Irradiated Fuel Examination Laboratory.

We would also like to thank the staff of the Nuclear Analytical Chemistry and Isotopics Laboratory for their work analyzing the sister rod fission gas samples and defueling the sister rod specimens. Particular thanks go to Jeff Delashmitt and Doug Canaan for their efforts. Thanks also to Ben Rothrock and his efforts on behalf of the sister rod project.

Thanks go to Tracy Binger and Mark Walls for their quick support when radiation protection coverage is required.

This page is intentionally blank

## CONTENTS

SUMMARY .....	D-iii
ACKNOWLEDGMENTS .....	D-v
CONTENTS.....	D-vii
LIST OF FIGURES .....	D-ix
LIST OF TABLES .....	D-xi
REVISION HISTORY .....	D-xiii
ACRONYMS .....	D-xv
D-1. Fission Gas Sample Isotopic Composition and Calculated Fission Gas Release .....	D-1
D-2. Fuel Burnup and Isotopic Measurements .....	D-6
D2.1 Separation and Measurement Techniques .....	D-6
D2.2 Calculation of Estimated Burnup .....	D-8
D2.3 Results .....	D-9
REFERENCES .....	D-11

This page is intentionally blank

## LIST OF FIGURES

Figure D-1. The sample inlet manifold (left) coupled to an OmniStar GSD 320 residual gas analyzer (RGA) (right). .....	D-1
--	-----

This page is intentionally left blank.

## LIST OF TABLES

Table DS-1. DE Status.....	D-iii
Table D-1. Ion source and focus lens tune parameters for the RGA analyzer.....	D-2
Table D-2. Matheson certified verification standards used for sister rod fission gas analysis. ....	D-3
Table D-3. Sister rod gas sample measured elemental composition, mole%* .....	D-3
Table D-4. Fission gas isotope ratios, atom% ratio .....	D-4
Table D-5. Fission gas isotopic composition, atom %* .....	D-5
Table D-6 List of specimens Selected for fuel burnup and isotopic measurements .....	D-6
Table D-7. Isotopes to be analyzed using high precision protocols.....	D-7
Table D-8. Chemical Isotopic Analysis (Burnup Only) of Sister Rod Specimens .....	D-10
Table D-9. Estimated Burnup for Sister Rod Specimens.....	D-10

This page is intentionally left blank.

**REVISION HISTORY**

<b>Date</b>	<b>Changes</b>
3/29/2019	Initial release
9/27/2019	Revised to include additional data and incorporate comments from the previously released report.
10/29/2020	The detailed information supporting the sister rod testing was moved into appendices to retain the information but reduce the amount of detail in the main report.
11/30/2020	The document numbering was revised to reflect its M2 status and the date was changed.

This page is intentionally left blank.

## ACRONYMS

CRM: certified reference material

DE	destructive examination
DOE	US Department of Energy
EPRI	Electric Power Research Institute
FIMA	fissions per initial metal atom
FHT	full-length fuel rod heat treatment
FY	fiscal year
GUM	Guide to the expression of uncertainty in measurements
HBU	high burnup
HP-RCA	High Precision Radiochemical analysis
ID	Isotope Dilution
IFEL	Irradiated Fuels Examination Laboratory
IHM	Initial Heavy Metal
IRMM	Institute for Reference Materials and Measurements (renamed as part of the European science hub to the joint research center at Geel (JRC-Geel))
LT	low tin
MC-ICPMS	Multi-collector Inductively Coupled Plasma Mass Spectrometer
MID	multiple ion detection
NBL	New Brunswick Laboratory
NE	Office of Nuclear Energy
NIST	National Institute of Standards and Technology
NDE	nondestructive examination
ORNL	Oak Ridge National Laboratory
PWR	pressurized water reactor
REDC	Radiochemical Engineering Development Center
RGA	residual gas analyzer
RPC	Research Project Cask
SEM	secondary electron multiplier
SFWD	Spent Fuel and Waste Disposition
SFWST	Spent Fuel and Waste Science and Technology
SNF	spent nuclear fuel
WRM	working reference material

This page is intentionally left blank.

## D-1. Fission Gas Sample Isotopic Composition and Calculated Fission Gas Release

Before the system evacuation for the void volume measurement, a fission gas sample was taken from each punctured sister rod. Gaseous fission products evolve in all uranium dioxide ( $\text{UO}_2$ ) nuclear fuel pellets at all axial elevations during reactor operation [D-1.] . They are located near the site of the fission, within the fuel grains, at a grain boundary, or at free surfaces on the pellet. The gaseous fission products form small bubbles within the pellet since the Xe and Kr gases produced are virtually insoluble in  $\text{UO}_2$ . Although much of the fission gas remains trapped within the fuel pellet microstructure as porosity, a fraction of the fission gas is released to the interior void volume of the fuel rod and contributes to an increase in the internal pressure of the fuel rod [D-2.] [D-3.] .

According to EPRI, less than 5% of the fission gas produced in the pellet stack during normal operation is released to the void volume of the rod [D-4.] . The quantity of fission gas released from the pellet to the void volume of the rod during reactor operation has been the topic of much study because the gross rod pressure and localized rod pressure are important to rod performance during reactor transients such as loss-of-coolant accidents and reactivity-initiated accidents. The percentage of fission gas released is calculated as the moles of fission gas in the rod void volume divided by the total calculated fission gas produced during operation.

The eight sister rod samples were analyzed by the Oak Ridge National Laboratory (ORNL) Nuclear Analytical Chemistry and Isotopic Laboratories Group. Fission gas isotopic (in atom %) and concentration (in mole%) determinations were made using an OmniStar GSD 320 residual gas analyzer (RGA) analyzer coupled to a sample manifold located within a radiological fume hood (Figure D-1). The ion source and focus lens tune parameters are listed in Table D-1.



**Figure D-1. The sample inlet manifold (left) coupled to an OmniStar GSD 320 residual gas analyzer (RGA) (right).**

**Table D-1. Ion source and focus lens tune parameters for the RGA analyzer.**

Parameter	Setting
Filament	No 1
Active set	Set 1
Emission current	0.50 A
Protection current	3.50 A
R <sub>f</sub> -polarity	positive
Ion reference	150.00 V
Cathode	-50.00 V
Focus	-3.00 V
Field axis	-5.00 V
Extraction	-45.00 V

The RGA analyzer is equipped with two detectors, a Faraday detector that is used for all fission gas concentration measurements and a secondary electron multiplier (SEM) that can be used for low parts per million gas concentration measurements. In general, the Faraday detector signal is more stable and offers slightly higher precision isotope ratio measurements. The SEM was not used for sister rod concentration measurements, but it was used to confirm the Faraday isotopic measurements. Gas samples are scanned from 1 to 150 atomic mass unit (amu) with either detector prior to starting the isotopic and mole percent determinations. These analog scans are performed to detect potential unknown gasses present in samples, and to have a record of detected peak shape and peak resolution. The natural (using a gas standard) or fission isotopes (unknown samples) are measured using a multiple ion detection (MID) data acquisition program. Gas concentrations are then measured by quantification of  $^{84}\text{Kr}$ ,  $^{132}\text{Xe}$ , and  $^4\text{He}$  isotopes using a multiple concentration determination program.

Quality control measures involved the analysis of two certified gas standard mixtures with independent production lot numbers. The primary standard is employed for calibration of the OmniStar GSD 320 RGA and the secondary standard is analyzed for calibration verifications and continuing instrument performance. The setup and calibration process of the OmniStar GSD 320 Gas Analysis (RGA) system involves an offset calibration with the evacuated sample manifold pressure controlled at 4–5 Torr. A mass scale adjustment and RGA ion source tuning are then performed with the certified mixed gas standard held at ~30 Torr in the sample manifold. Mass scale tuning, which is required for performance optimization, is performed when necessary. The final stage in the RGA setup is a second mass scale adjustment using a fine mass calibration setting and the same standard gas. The result of the calibration is verified with the secondary mixed gas standard held at the same pressure.

To measure the sister rod fission gas samples, the sample bottle used in the rod puncture sequence (Appendix C) is attached to the evacuated sample manifold for the measurement sequence. The known volume of the sample manifold is 18.4 mL, and the volume and pressure of each sister rod gas sample is known. The sample gas expands into the manifold until gas pressure reaches equilibrium. The sample manifold pressure is recorded and then reduced to ~30 Torr before injecting the sample into the RGA for analysis. Argon gas (99.999% purity) is used to flush the sample manifold between gas analyses and for verification testing. NIST traceable mixed gas standards from independent lots are used for calibration and verification testing, and these contain a gas mixture that is 5.00 mol% Kr, 5.00 mol% Xe, 1.05 mol% O, 3.96 mol% N, and the remaining 85.00 mol% is He. The individual gas concentrations in the mixed gas standards are certified with a blend tolerance and certified accuracy of a relative  $\pm 2\%$ , and the certified test gas standards used are listed in Table D-2.

**Table D-2. Matheson certified verification standards used for sister rod fission gas analysis.**

Calibration-	Lot No:	1028601078B	Exp.	4/20/2021
Verification-	Lot No:	1027603245B	Exp.	12/8/2020

Measured gas concentrations for the eight sister rod gas samples collected during the rod puncture operation (see Section 7.1) are provided in Table D-3. Six of the sister rod samples were measured as many as three times on non-consecutive days and the determined fission gas concentrations were averaged for those samples. The concentrations measured were determined by linear regression monitoring  $^{84}\text{Kr}$  and  $^{132}\text{Xe}$ , which are naturally occurring isotopes present at 56.99 and 26.91 atom%, respectively. The isotopic concentration in the sister rod samples was determined by measuring the current responses corresponding to the  $^{84}\text{Kr}$  and  $^{132}\text{Xe}$  isotopes and comparing those with the current response of the known concentration calibration standards. The total uncertainty values reported are the combined uncertainties of the duplicate measurements at a 95% level of confidence. The number of digits in the reported mole% and their uncertainties are provided for information and are not intended to convey a significant degree of reliability for the long-term performance of precision for the measurement.

**Table D-3. Sister rod gas sample measured elemental composition, mole%\***

Detected gas**	Sample ID											
	30AK09 (M5) SR-Gr-02				30AD05 (M5) SR-Gr-05			30AE14 (M5, heat treated) SR-Gr-06			3A1F05 (LT Zirc-4) SR-Gr-04	
Kr	1.60	±	0.15	1.41	±	0.19	1.45	±	0.22	1.97	±	0.25
Xe	15.31	±	1.33	14.10	±	0.70	14.11	±	1.49	18.46	±	1.72
He***	83.09	±	1.10	84.49	±	0.59	84.44	±	1.62	79.57	±	1.83

  

Detected gas**	Sample ID											
	6U3K09 (ZIRLO) SR-Gr-01				3D8E14 (ZIRLO) SR-Gr-03			3F9N05 (ZIRLO, heat treated) SR-Gr-07			F35P17 (Zirc-4, heat treated) SR-Gr-08	
Kr	1.11	±	0.10	2.36	±	0.30	2.23	±	0.30	1.93	±	0.25
Xe	10.45	±	1.47	22.44	±	1.41	20.08	±	2.01	19.87	±	1.99
He***	88.44	±	1.41	75.20	±	1.41	77.69	±	1.62	78.20	±	1.62

\* Reported uncertainties are the total combined uncertainties at the 95% level of confidence. Two decimal places are provided in the reported values for information only and are not intended to imply a significant degree of reliability. The precision contribution for samples 01-06 was the standard deviation of the values measured in August 2018 and September 2018. Because only a single data set was measured for samples 07 and 08, for conservatism, the precision contribution to the total uncertainty for those data was taken as the worst-case scenario observed for samples 01-06.

\*\* Some residual air present in the sampling system was detected and the resulting O and N content has been neglected when determining the fission gas component percentages and fission gas release in the fuel rod.

\*\*\*The measured He includes the pre-pressurization He and any He produced as fission/decay products.

Isotopic data, reported in Table D-4 and Table D-5, were determined using the ion currents acquired from the MID run and includes both natural and fission product Kr and Xe isotopes. The isotopic composition of natural Kr and Xe in both primary and secondary standards were evaluated and compared with isotopic data sourced from the National Nuclear Data Center [D-5.] to ensure analytical accuracy. The uncertainties reported in Table D-4 are a single standard deviation from the calculated mean, where  $n = 2-4$ , except for SR-Gr-07 and -08 in which only one measurement was completed. For those two samples, the averaged observed standard deviation for the other six samples was applied. The uncertainties reported in Table D-5

are  $2\sigma$  standard deviations from the calculated mean, where  $n = 2-4$ , except for SR-Gr-07 and -08 in which only one measurement was completed. For those two samples, the largest observed uncertainty from the other six samples was applied. The number of digits in the reported atom% and their uncertainties are provided for information and are not intended to convey a significant degree of reliability. Some small bias in the measurement data was observed but was within the measurement uncertainty; no corrections were applied. Additionally, three independent measurements of selected sister rod samples were performed on three different dates between August 2018 and February 2019. The Faraday detector was used to generate most of the isotopic data; however, the SEM was also used for three of the fission gas samples in February 2019 to verify the precision of the Faraday measurements. The results of the three SEM measurements were incorporated into the uncertainty calculations for the measured samples.

**Table D-4. Fission gas isotope ratios, atom% ratio**

Sample Isotope Ratio	30AK09 (M5) SP-Gr-02*	30AD05 (M5) SP-Gr-05*	30AE14 (M5, heat treated) SP-Gr-06*	3A1F05 (LT Zirc-4) SP-Gr-04*
$^{82}\text{Kr}/^{84}\text{Kr}$	0.021 +/- 0.009	0.018 +/- 0.007	0.019 +/- 0.008	0.023 +/- 0.003
$^{83}\text{Kr}/^{84}\text{Kr}$	0.322 +/- 0.009	0.307 +/- 0.008	0.318 +/- 0.010	0.310 +/- 0.002
$^{85}\text{Kr}/^{84}\text{Kr}^{***}$	0.121 +/- 0.007	0.121 +/- 0.008	0.123 +/- 0.009	0.059 +/- 0.003
$^{86}\text{Kr}/^{84}\text{Kr}$	1.533 +/- 0.014	1.525 +/- 0.009	1.535 +/- 0.011	1.512 +/- 0.000
$^{128}\text{Xe}/^{132}\text{Xe}$	0.005 +/- 0.003	0.005 +/- 0.003	0.006 +/- 0.003	0.007 +/- 0.004
$^{130}\text{Xe}/^{132}\text{Xe}$	0.012 +/- 0.002	0.012 +/- 0.002	0.013 +/- 0.001	0.015 +/- 0.001
$^{131}\text{Xe}/^{132}\text{Xe}$	0.278 +/- 0.004	0.289 +/- 0.003	0.296 +/- 0.002	0.293 +/- 0.004
$^{134}\text{Xe}/^{132}\text{Xe}$	1.178 +/- 0.006	1.173 +/- 0.022	1.178 +/- 0.011	1.165 +/- 0.033
$^{136}\text{Xe}/^{132}\text{Xe}$	1.689 +/- 0.029	1.661 +/- 0.055	1.654 +/- 0.029	1.647 +/- 0.080
Sample Isotope Ratio	6U3K09 (ZIRLO) SP-Gr-01*	3D8E14 (ZIRLO) SP-Gr-03*	3F9N05 (ZIRLO, heat treated) SP-Gr-07**	F35P17 (Zirc-4, heat treated) SP-Gr-08**
$^{82}\text{Kr}/^{84}\text{Kr}$	0.016 +/- 0.005	0.022 +/- 0.004	0.033 +/- 0.010	0.034 +/- 0.011
$^{83}\text{Kr}/^{84}\text{Kr}$	0.311 +/- 0.005	0.277 +/- 0.001	0.310 +/- 0.006	0.278 +/- 0.005
$^{85}\text{Kr}/^{84}\text{Kr}^{***}$	0.100 +/- 0.004	0.073 +/- 0.004	0.088 +/- 0.005	0.050 +/- 0.003
$^{86}\text{Kr}/^{84}\text{Kr}$	1.537 +/- 0.008	1.474 +/- 0.007	1.530 +/- 0.008	1.469 +/- 0.008
$^{128}\text{Xe}/^{132}\text{Xe}$	0.005 +/- 0.003	0.007 +/- 0.004	0.011 +/- 0.006	0.011 +/- 0.006
$^{130}\text{Xe}/^{132}\text{Xe}$	0.012 +/- 0.002	0.015 +/- 0.002	0.019 +/- 0.002	0.018 +/- 0.002
$^{131}\text{Xe}/^{132}\text{Xe}$	0.278 +/- 0.004	0.254 +/- 0.005	0.290 +/- 0.004	0.249 +/- 0.003
$^{134}\text{Xe}/^{132}\text{Xe}$	1.178 +/- 0.006	1.126 +/- 0.033	1.156 +/- 0.018	1.094 +/- 0.017
$^{136}\text{Xe}/^{132}\text{Xe}$	1.689 +/- 0.029	1.582 +/- 0.083	1.600 +/- 0.048	1.545 +/- 0.046

\* Uncertainty for samples defined as a  $1\sigma$  external standard deviation of the replicate analyses (for 01, 03, and 04,  $n = 2$ ; for 02, 05, and 06,  $n = 4$ ).

\*\* For SP-Gr-07 and -08, only one replicate was performed; the assigned uncertainties are the averages of the other six samples.

\*\*\*  $^{85}\text{Kr}$  was decay-corrected to February 2019 in each case.

Table D-5. Fission gas isotopic composition, atom %\*

Sample Isotope	30AK09 (M5) SP-Gr-02*	30AD05 (M5) SP-Gr-05*	30AE14 (M5, heat treated) SP-Gr-06*	3A1F05 (LT Zirc-4) SP-Gr-04*
<sup>82</sup> Kr	0.69 ± 0.34	0.59 ± 0.30	0.64 ± 0.32	0.81 ± 0.40
<sup>83</sup> Kr	10.73 ± 0.32	10.34 ± 0.31	10.62 ± 0.32	10.67 ± 0.32
<sup>84</sup> Kr	33.37 ± 0.67	33.65 ± 0.67	33.39 ± 0.67	34.43 ± 0.69
<sup>85</sup> Kr****	4.04 ± 0.20	4.08 ± 0.20	4.12 ± 0.21	2.03 ± 0.10
<sup>86</sup> Kr	51.17 ± 0.51	51.33 ± 0.51	51.24 ± 0.51	52.06 ± 0.52
<sup>128</sup> Xe	0.14 ± 0.07	0.13 ± 0.07	0.17 ± 0.09	0.16 ± 0.08
<sup>129</sup> Xe	< 0.05	< 0.05	< 0.05	< 0.05
<sup>130</sup> Xe	0.31 ± 0.16	0.30 ± 0.15	0.39 ± 0.19	0.36 ± 0.18
<sup>131</sup> Xe	7.09 ± 0.35	6.98 ± 0.35	6.38 ± 0.32	7.11 ± 0.36
<sup>132</sup> Xe	23.97 ± 0.48	24.16 ± 0.48	25.14 ± 0.50	24.24 ± 0.48
<sup>134</sup> Xe	28.39 ± 0.57	28.33 ± 0.57	28.30 ± 0.57	28.23 ± 0.56
<sup>136</sup> Xe	40.15 ± 0.40	40.10 ± 0.40	39.74 ± 0.40	39.90 ± 0.40
Sample Isotope	6U3K09 (ZIRLO) SP-Gr-01*	3D8E14 (ZIRLO) SP-Gr-03*	3F9N05 (ZIRLO, heat treated) SP-Gr-07**	F35P17 (Zirc-4, heat treated) SP-Gr-08**
<sup>82</sup> Kr	0.56 ± 0.28	0.77 ± 0.39	1.10 ± 0.55	1.20 ± 0.60
<sup>83</sup> Kr	10.49 ± 0.31	9.75 ± 0.29	10.45 ± 0.31	9.83 ± 0.29
<sup>84</sup> Kr	33.74 ± 0.67	35.14 ± 0.70	33.74 ± 0.67	35.30 ± 0.71
<sup>85</sup> Kr****	3.36 ± 0.17	2.56 ± 0.13	3.06 ± 0.15	1.81 ± 0.09
<sup>86</sup> Kr	51.85 ± 0.52	51.78 ± 0.52	51.63 ± 0.52	51.86 ± 0.52
<sup>128</sup> Xe	0.12 ± 0.06	0.17 ± 0.09	0.27 ± 0.20	0.28 ± 0.21
<sup>129</sup> Xe	< 0.05	< 0.05	< 0.05	< 0.05
<sup>130</sup> Xe	0.29 ± 0.15	0.39 ± 0.19	0.46 ± 0.23	0.45 ± 0.23
<sup>131</sup> Xe	6.68 ± 0.33	6.38 ± 0.32	7.12 ± 0.36	6.36 ± 0.32
<sup>132</sup> Xe	24.03 ± 0.48	25.14 ± 0.50	24.54 ± 0.49	25.54 ± 0.51
<sup>134</sup> Xe	28.31 ± 0.57	28.30 ± 0.57	28.36 ± 0.57	27.92 ± 0.56
<sup>136</sup> Xe	40.57 ± 0.41	39.74 ± 0.40	39.25 ± 0.39	39.44 ± 0.39

\* Reported numerical uncertainties are the 2σ external standard deviation of all duplicate analyses. The last digit in the measurements and uncertainties is provided for information and is not intended to convey a significant degree of reliability. The accuracy of the analysis was confirmed using a NIST traceable standard, and a bias correction did not measurably alter the data within the uncertainty of the 2σ standard deviation.

\*\* SR-Gr-02, -05, and -06 also incorporate the uncertainty between two different modes of mass analysis, namely SEM and Faraday.

\*\*\* For SR-Gr-07 and -08, only one replicate was analyzed; for conservatism, the uncertainty attributed to those data was taken as the worst-case scenario observed for samples 01–06.

\*\*\*\* <sup>85</sup>Kr was decay-corrected to February 2019 in each case.

## D-2. Fuel Burnup and Isotopic Measurements

Eleven PWR samples were selected from the sister rods, as listed in Table D-6. Three will be analyzed for chemical determination of burnup (called *burnup only*) and eight will be analyzed using high precision protocols for the isotopes listed in Table D-7 (called *full-isotopic analysis*). The three samples selected for burnup only measurements have operator-reported burnups of 41, 57, and 64 GWd/MTU. The eight samples planned for full-isotopic analysis cover a range of burnups from 53 to 64 GWd/MTU.

**Table D-6 List of specimens Selected for fuel burnup and isotopic measurements**

Rod	Originating segment elevation range (mm)		Reported burnup (GWd/MTU)	Analysis performed	Sample length (mm)	Status
6U3K09	3506	3525	43	Burnup only	12.4	Complete
3D8E14	3206	3225	59	Burnup only	8.6	Complete
3D8E14	700	719	61	Burnup only	14.9	Complete
3A1F05	2383	2402	53	Full isotopic	9.8	To be dissolved
3F9N05	2300	2329	57	Full isotopic	10.3	To be dissolved
3F9N05	2863	2882	58	Full isotopic	9.1	To be dissolved
30AD05	1280	1299	59	Full isotopic	11.7	To be dissolved
30AD05	2410	2429	59	Full isotopic	9.2	To be dissolved
30AE14	2675	2694	61	Full isotopic	8.9	Dissolved
3D8E14	2303	2322	62	Full isotopic	12.4	To be dissolved
3D8E14	2655	2674	64	Full isotopic	8.3	To be dissolved
<b>Total number of specimens</b>					<b>11</b>	

### D2.1 Separation and Measurement Techniques

Chemical separation is achieved using ion chromatography using oxalic acid, diglycolic acid, and hydrochloric acid. A method called RAPID (Rapid Analysis of Post-Irradiation Debris) [D-6,D-7,D-8,D-9,D-10], originally developed for the Defense Threat Reduction Agency (DTRA), has been modified for the successful separation of and isolation of all activation and fission isotopes, including  $^{125}\text{Sb}$ ,  $^{95}\text{Mo}$ ,  $^{99}\text{Tc}$ , and  $^{109}\text{Ag}$ . RAPID, when used in combination with isotope dilution mass spectrometry (IDMS) can yield direct isotopic compositions and concentrations with uncertainties as low as 2% at the 95% confidence level using a single detector inductively coupled plasma mass spectrometer (ICP-MS). When coupled with elemental isolation via fraction collection followed by analysis using a multi-collector ICP-MS isotopic the IDMS-RAPID protocol can yield isotopic concentrations with uncertainties as low as 0.5% at the 95% confidence level.

ORNL has developed and qualified a comprehensive analytical protocol for high precision radiochemical analysis (HP-RCA) of key fission product and actinide isotopes in select specimens from the HBU fuels. Table D-7 provides a summary of the improvements in uncertainty afforded by the high precision approach as compared with the previous analysis approach and lists the isotopes to be analyzed. A key to the HP-RCA is the state-of-the-art MC-ICPMS system installed for use with radioactive samples. This instrument is capable of measuring isotope ratios with relative uncertainties on the order of 0.1%. A comprehensive

Sample Analysis Plan for the HP-RCA measurements has been developed [D-11.] and a working reference material (WRM) has been prepared using NIST and NBL certified reference materials (CRM). For the HP-RCA, spikes were created using enriched isotopes procured from the DOE's Isotope Business Offices with traceability established to NIST standards. The WRM will be processed in parallel with the fuels and used as a laboratory control sample (LCS) to document performance of the analytical protocol. A full list of standards and controls can be found in Appendix D in ORNL/TM-202/1657 [D-11.] .

Measurement protocols with GUM-compliant<sup>1</sup> calculations to report total uncertainties for the results were established for all nuclides listed in Table D-7.

ORNL is unable to measure <sup>14</sup>C; <sup>36</sup>Cl, <sup>221</sup>Fr with current analytical capabilities.

**Table D-7. Isotopes to be analyzed using high precision protocols.**

Isotope	Separation and detection method	Relative uncertainty single detector ICPMS, 2 $\sigma$ * (prior best practice)	Relative uncertainty HP-RCA test method, 2 $\sigma$ *
<sup>79</sup> Se	RAPID-IDMS with a <sup>77</sup> Se spike	2%	0.5%
<sup>95</sup> Mo	RAPID-IDMS with a <sup>94</sup> Mo spike	-	0.5%
<sup>90</sup> Sr	RAPID-IDMS with a <sup>86</sup> Sr spike	5%	0.5%
<sup>99</sup> Tc	Standard addition using <sup>99</sup> Tc	-	2%
<sup>101</sup> Ru	RAPID-IDMS using a <sup>99</sup> Ru spike	2%	0.5%
<sup>106</sup> Ru			0.5%
<sup>103</sup> Rh	Standard Addition	5%	-
<sup>109</sup> Ag	Rapid-IDMS using a <sup>107</sup> Ag spike	-	1%
<sup>125</sup> Sb	RAPID-IDMS using a <sup>123</sup> Sb spike	2%	0.5%
<sup>133</sup> Cs	RAPID-IDMS using a <sup>133</sup> Cs spike	2%	0.5%
<sup>134</sup> Cs			0.5%
<sup>135</sup> Cs			0.5%
<sup>137</sup> Cs**			0.5%
<sup>143</sup> Nd	RAPID-IDMS using a <sup>150</sup> Nd spike	2%	0.5%
<sup>145</sup> Nd			0.5%
<sup>146</sup> Nd			0.5%
<sup>148</sup> Nd**			0.5%
<sup>144</sup> Nd			0.5%
<sup>144</sup> Ce	RAPID-IDMS using a <sup>140</sup> Ce spike	2%	1.0%
<sup>147</sup> Pm	RAPID-IDMS using <sup>149</sup> Sm and a semi-quantitative approach	3%	-
<sup>147</sup> Sm	RAPID-IDMS using a <sup>149</sup> Sm spike	2%	0.5%
<sup>149</sup> Sm			0.5%
<sup>150</sup> Sm			0.5%

<sup>1</sup> ISO/IEC GUIDE 98-3:2008, Guide to the expression of uncertainty in measurement.

Isotope	Separation and detection method	Relative uncertainty single detector ICPMS, 2 $\sigma^*$ (prior best practice)	Relative uncertainty HP-RCA test method, 2 $\sigma^*$
<sup>151</sup> Sm			0.5%
<sup>152</sup> Sm			0.5%
<sup>151</sup> Eu	RAPID-IDMS using a <sup>151</sup> Eu spike	2%	0.5%
<sup>153</sup> Eu			0.5%
<sup>154</sup> Eu			0.5%
<sup>155</sup> Eu			0.5%
<sup>155</sup> Gd	RAPID-IDMS using <sup>155</sup> Gd	2%	0.5%
<sup>227</sup> Ac	Weighted dilution gamma	5%	-
<sup>234</sup> U	Davies-Grey titration or IDMS with <sup>233</sup> U	0.5% or 2%	0.1%
<sup>235</sup> U**			0.1%
<sup>236</sup> U			0.1%
<sup>238</sup> U**			0.1%
<sup>237</sup> Np	RAPID using external calibration	5%	-
<sup>238</sup> Pu	RAPID-IDMS using CRM-130, a <sup>242</sup> Pu STD	2%	0.1%
<sup>239</sup> Pu**			0.1%
<sup>240</sup> Pu**			0.1%
<sup>241</sup> Pu**			0.1%
<sup>242</sup> Pu			0.1%
<sup>241</sup> Am	RAPID-IDMS using an <sup>243</sup> Am , <i>verification of 241 using gamma</i>	2%	0.1%
<sup>242m</sup> Am			0.5%
<sup>243</sup> Am			0.1%
<sup>242</sup> Cm	RAPID IDMS with in-house certified Cm-WRM, or using <sup>243</sup> Am and a semiquantitative approach	2%	1.0%
<sup>243</sup> Cm			1.0%
<sup>244</sup> Cm			1.0%
<sup>245</sup> Cm			1.0%
<sup>246</sup> Cm			1.0%
<sup>247</sup> Cm			1.0%

\*Based on the isotopic concentration of 1  $\mu\text{g/g}$  fuel

\*\*Isotopes required for the determination of chemical burnup

## D2.2 Calculation of Estimated Burnup

Burnup is estimated based on the measurement data using ASTM-321E standard methods [D-12.] , relying on <sup>148</sup>Nd as a burnup indicator. The sample's burnup, B in GWd/MTU, is determined as the product of the estimated number of fissions per initial metal atom (FIMA) and a unit conversion factor C that relates FIMA% to GWd/MTU burnup units:

$$B = C \times FIMA\% \quad (\text{D-1})$$

The recommended value for C is  $9.6 \pm 0.3$  GWd/MTU [D-12.] .

FIMA% was determined based on the measurement data for  $^{148}\text{Nd}$  and uranium and plutonium isotopes:

$$FIMA\% = \frac{{}^{148}\text{Nd}/Y_{eff}}{(U+Pu+{}^{148}\text{Nd}/Y_{eff})} \times 100 \quad (\text{D-2})$$

where

${}^{148}\text{Nd}$  is number of  ${}^{148}\text{Nd}$  atoms measured in the sample,

$Y_{eff}$  is effective fission yield for  ${}^{148}\text{Nd}$ ,

$U$  is number of uranium atoms measured in the sample, and

$Pu$  is number of plutonium atoms measured in the sample.

The estimated burnup uncertainty is inferred by error propagation in (D-1), assuming all variables involved are independent:

$$\sigma_B^2 = \sum_{i=1}^N \left( \frac{\delta B}{\delta x_i} \right)^2 \sigma_i^2 \quad (\text{D-3})$$

The uncertainty used for the number of  ${}^{148}\text{Nd}$  atoms measured in the sample combined the measurement uncertainty for the isotope with an estimated uncertainty that corrects for radiative captures in  ${}^{147}\text{Nd}$  and  ${}^{148}\text{Nd}$  [D-8]. This correction is expected to be small (<0.5%).

## D2.3 Results

Chemical determination of burnup per the established ASTM method [D-13.] has been completed for the three selected fuel specimens. A drop of sample 3D8E14-3206-3225 was lost during filtering of the final digested solution in the hot cell. At the time of loss, the solution was homogeneous and therefore the ratio of Nd-148 burnup indicator to uranium and plutonium was not compromised. For this sample burnup was calculated using total atom ratios in the final solution vs. the pellet as is convention. Performing the calculation in this manner does not affect the results and they are considered accurate.

The isotope measurement data used for burnup determination for the 3 samples are provided in terms of g/g initial heavy metal (IHM) in Table D-8; the uncertainty is  $1\sigma$ .

The estimated sample burnups and the FIMA% values based on the measurement data are presented in Table D-8. A value of 1.706% [D-13.] was used for the effective fission yield of  ${}^{148}\text{Nd}$ . The quantified  $1\sigma$  burnup uncertainties for U and Pu are consistent with the measurement data in Table D-8. The uncertainty used for these calculations was conservatively assumed as 0.5% [D-14.] .

Table D-9 also compares the measurement-based burnup estimate with the operator-reported burnup. There is good agreement, with a relative difference on the order of 5%. 3D8E14-3206-3225 and 6U3K09-3506-3525 are both from the steep burnup gradient region at the top of the fuel rod, whereas, sample 3D8E14-700-719 was cut from the flat burnup region of the rod. The samples from the steep gradient region are likely to have a higher uncertainty associated with the operator-reported burnup that is related to the fidelity of the simulations, and this is consistent with the differential from the measured burnup, with the steep gradient samples having a slightly higher differential from predictions.

Additionally, as a proof of concept, the Nd fractions from the three burnup samples were analyzed using HP-RCA. A comparison of the two datasets shows an improvement in relative precision for elemental Nd from +/- 3 % to +/- 0.5 % and its isotopic abundances from +/- 0.08 % to +/- 0.005 % ( $2\sigma$ ).

Table D-8. Chemical Isotopic Analysis (Burnup Only) of Sister Rod Specimens

Project ID	3D8E14-700-719		3D8E14-3206-3225		6U3K09-3506-3525	
Lab ID	TAL SR-719		TAL SR-3225		TAL SR-3525	
Specimen weight (g)	6.648		4.532		6.707	
Units	g/gIHM	1 $\sigma$ uncertainty	g/gU	1 $\sigma$ uncertainty	g/gIHM	1 $\sigma$ uncertainty
Nd (isotopics over six runs)	7.526E-03	3.919E-05	7.724E-03	3.987E-05	5.237E-03	2.755E-05
<sup>124</sup> Nd wt %	0.8687%	0.0080%	0.7634%	0.0034%	0.4903%	0.0039%
<sup>143</sup> Nd wt %	15.1330%	0.0017%	16.3629%	0.0011%	20.1298%	0.0016%
<sup>144</sup> Nd wt %	36.4023%	0.0034%	35.2676%	0.0019%	32.3057%	0.0021%
<sup>145</sup> Nd wt %	14.9920%	0.0016%	15.2970%	0.0010%	16.5945%	0.0014%
<sup>146</sup> Nd wt %	18.7087%	0.0025%	18.4323%	0.0014%	17.1489%	0.0016%
<sup>148</sup> Nd wt %	9.2956%	0.0019%	9.3012%	0.0016%	9.0248%	0.0016%
<sup>150</sup> Nd wt %	4.5996%	0.0017%	4.5756%	0.0016%	4.3059%	0.0015%
Units	g/gIHM	1 $\sigma$ uncertainty	g/gSoln	1 $\sigma$ uncertainty	g/gIHM	1 $\sigma$ uncertainty
U	9.150E-01	3.235E-03	3.424E-03	0.342E-05	9.418E-01	1.561E-03
<sup>233</sup> U wt %	0.0010%	NA	0.0010%	NA	0.0010%	NA
<sup>234</sup> U wt %	0.0248%	0.0002%	0.0254%	0.0002%	0.0272%	0.0003%
<sup>235</sup> U wt %	0.4368%	0.0005%	0.5822%	0.0007%	1.3359%	0.0016%
<sup>236</sup> U wt %	0.6425%	0.0066%	0.6370%	0.0066%	0.5944%	0.0061%
<sup>238</sup> U wt %	98.8958%	0.0066%	98.7554%	0.0066%	98.0425%	0.0063%
Units	g/gIHM	1 $\sigma$ uncertainty	g/gU	1 $\sigma$ uncertainty	g/gU	1 $\sigma$ uncertainty
Pu (isotopics over six runs)	1.119E-02	1.976E-04	1.276E-02	4.890E-04	1.028E-02	7.288E-05
<sup>238</sup> Pu wt %	3.4261%	0.0598%	3.6434%	0.2286%	2.1232%	0.0597%
<sup>239</sup> Pu wt %	50.9118%	0.1726%	51.7021%	0.9671%	59.1770%	0.4432%
<sup>240</sup> Pu wt %	27.8080%	0.1372%	27.3843%	0.8421%	24.6114%	0.4178%
<sup>241</sup> Pu wt %	6.3293%	0.1139%	6.5196%	0.2642%	8.3739%	0.1013%
<sup>242</sup> Pu wt %	11.5248%	0.1091%	10.7505%	0.3482%	5.7145%	0.0265%

Table D-9. Estimated Burnup for Sister Rod Specimens

Project ID	3D8E14-700-719	3D8E14-3206-3225	6U3K09-3506-3525
Lab ID	TAL SR-719	TAL SR-3225	TAL SR-3525
Measured FIMA (%)	6.651	6.239	4.473
Measured burnup (GWd/MTU)	63.849	59.895	42.940
Measured burnup uncertainty, 1 $\sigma$ (%)	0.9	0.8	0.7
Operator-estimated burnup (GWd/MTU)	63.564	56.779	40.658
Measured/operator burnup ratio	1.004	1.055	1.056

## REFERENCES

- [D-1.] S. Katco, "Fission-Product Yields from Neutron-Induced Fission," *Nucleonics*. **18** (11):201–208 (1960).
- [D-2.] M. V. Speight, "A Calculation on the Migration of Fission Gas in Material Exhibiting Precipitation and Re-solution of Gas Atoms Under Irradiation," *Nucl. Sci. Eng.* **37**, 180–185 (1969).
- [D-3.] J. A. Turnbull, "Distribution of Intragranular Fission Gas Bubbles in UO<sub>2</sub> During Irradiation," *J. Nucl. Mater.* **38** (2), 203 (1971).
- [D-4.] Topical Report on Reactivity Initiated Accident: Bases for RIA Fuel and Core Coolability Criteria, 3-1 – 3-9, EPRI, Palo Alto, California, 1002865 (2002).
- [D-5.] Laboratory BN National Nuclear Data Center. <http://www.nndc.bnl.gov/>.
- [D-6.] Roach BD, Fenske EK, Ilgner RH, et al. (2018). Development of a fast and efficient analytical technique for the isotopic analysis of fission and actinide elements in environmental matrices. *J Chromatogr A* 1–11 . doi: 10.1016/j.chroma.2018.12.029
- [D-7.] Roach BD, Hexel CR, Giaquinto JM (2017). ORNL TM-2017-694: Demonstration of a Rapid HPLC-ICPMS Direct Coupling Technique Using IDMS, Project Report Part I.
- [D-8.] Roach BD, Glasgow DC, Fenske EK, et al. (2017). ORNL TM-2017-697: Demonstration of a Rapid HPLC-ICPMS Direct Coupling Technique Using IDMS, Project Report Part II.
- [D-9.] Fenske EK, Roach BD, Glasgow DC, et al. (2019). Rapid measurements of U235 fission product isotope ratios with online high-pressure ion chromatography ICPMS protocol and comparison to isotopic depletion models. *J Radioanal Nucl Chem*.
- [D-10.] Roach BD, Fenske EK, Ilgner RH, et al. (2017). RAPID: Development of a fast and efficient analytical technique for the isotopic analysis of fission and actinide elements in environmental matrices. *J Chromatogr. A*
- [D-11.] ORNL/TM-2020/1657 Nuclear Analytical and Chemical Isotopics Laboratories Sample Analytical Plan for the Sister Rod Spent Fuel Specimens.
- [D-12.] American Society for Testing and Materials Standard Method E 321-96, "Standard Test Method for Atom Percent Fission in Uranium and Plutonium Fuel (Neodymium-148 Method)," (2005).
- [D-13.] I.C. Gauld, J. M. Giaquinto, J. S. Delashmitt, J. Hu, G. Ilas, T.J. Haverlock, C. Romano, " Re-evaluation of Spent Nuclear Fuel Assay Data for the Three Mile Island Unit 1 Reactor and Application to Code Validation ", *Annals of Nuclear Energy* vol.87, pp. 267-281 (2016) <https://doi.org/10.1016/j.anucene.2015.08.026>.
- [D-14.] K. Suyama and H. Mochizuki, Corrections to the <sup>148</sup>Nd method of evaluation of burnup for the PIE samples from Mihama-3 and Genkai-1 reactors, *Ann. Nucl. Energy* 33, pp 335-342 (2006).