

**ELECTRONICS PERSONAL DOSEMETER  
(EPD-N)  
TEST AND EVALUATION REPORT**

By  
Ayman A. Shourbaji  
Peter J. Chiaro  
John Dixon

## DOCUMENT AVAILABILITY

Reports produced after January 1, 1996, are generally available free via the U.S. Department of Energy (DOE) Information Bridge.

**Web site** <http://www.osti.gov/bridge>

Reports produced before January 1, 1996, may be purchased by members of the public from the following source.

National Technical Information Service  
5285 Port Royal Road  
Springfield, VA 22161  
**Telephone** 703-605-6000 (1-800-553-6847)  
**TDD** 703-487-4639  
**Fax** 703-605-6900  
**E-mail** [info@ntis.fedworld.gov](mailto:info@ntis.fedworld.gov)  
**Web site** <http://www.ntis.gov/support/ordernowabout.htm>

Reports are available to DOE employees, DOE contractors, Energy Technology Data Exchange (ETDE) representatives, and International Nuclear Information System (INIS) representatives from the following source.

Office of Scientific and Technical Information  
P.O. Box 62  
Oak Ridge, TN 37831  
**Telephone** 865-576-8401  
**Fax** 865-576-5728  
**E-mail** [reports@adonis.osti.gov](mailto:reports@adonis.osti.gov)  
**Web site** <http://www.osti.gov/contact.html>

## DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

## COPYRIGHT NOTICE

The submitted manuscript has been authored by a contractor of the U.S. Government under contract DE-AC05-00OR22725. Accordingly, the U.S. Government retains a paid-up, nonexclusive, irrevocable, worldwide license to publish or reproduce the published form of this contribution, prepare derivative works, distribute copies to the public, and perform publicly and display publicly, or allow others to do so, for U.S. Government purposes.

**ELECTRONICS PERSONAL DOSEMETER  
(EPD-N)  
TEST AND EVALUATION REPORT**

By  
Ayman A. Shourbaji  
Peter J. Chiaro  
John Dixon

Engineering Science and Technology Division

September 2002

Prepared by  
OAK RIDGE NATIONAL LABORATORY  
Oak Ridge, Tennessee 37831-6285  
managed by  
UT-BATTELLE, LLC  
for the  
U.S. DEPARTMENT OF ENERGY  
under contract DE-AC05-00OR22725

## Table of Contents

|  |   |
|--|---|
| Executive Summary .....  | 1 |
| 1. Introduction .....  | 2 |
| 2. Neutron Test .....  | 2 |
| 3. Gamma Test Using $^{137}\text{Cs}$ .....                              | 3 |
| 4. Gamma Test Using $^{60}\text{Co}$ .....                               | 3 |
| 5. X-ray Test.....   | 3 |
| Table 1. Gamma measurement using $^{137}\text{Cs}$ .....                 | 4 |
| Table 2. Gamma-Neutron cross talk measurement ( $^{137}\text{Cs}$ )..... | 4 |
| Table 3. Gamma test results for $^{60}\text{Co}$ .....                   | 4 |
| Table 4. X-ray measurement results .....                                 | 4 |
| Appendix A, EPD-N Published Specifications.....                          | 5 |

## **Executive Summary**

Three electronic personal dosimeters (EPD-N) manufactured by Siemens, serial numbers 0635, 0658, and 0683, were tested at the Radiation Calibration Laboratory for an evaluation of their response to neutron, gamma and x-ray radiation. Designed to provide real-time neutron and photon dosimetry, the EPD-N is capable of estimating and displaying neutron and gamma dose components separately for a range of energies from 50 keV to 7 MeV for photon and 0.025 eV to > 10 MeV for neutron.

All tests were conducted using the factory calibrations. A technical representative of the manufacturer indicated that site-specific calibrations are required as factory settings are calibrated for the lowest neutron energy limit of 0.025 eV. This raises concerns about the reliability of these devices in measuring neutrons when calibrations are made for a specific site radiological characterization then used at another site.

## 1. Introduction

Three electronic personal dosimeters (EPD-Ns) manufactured by Siemens (S/Ns: 0635, 0658, 0683) were tested at the Radiation Calibration Laboratory to evaluate their response to neutrons, gamma, and x-ray radiations. The EPD-N is designed to provide real-time neutron and photon dosimetry. It is capable of estimating and displaying neutron and gamma dose components separately for a wide range of energies (50 keV to 7 MeV for photon, and 0.025 eV to > 10 MeV for neutron). The manufacturer's published specifications are shown in Attachment A of this report. The purpose of this effort is to evaluate the performance of the EPD-N device as a candidate for personal dosimeter use at Oak Ridge National Laboratory.

Tests were performed using the factory calibration. All tests were conducted with the units mounted on a 40 cm x 40 cm x 15 cm Plexiglas phantom to simulate actual application conditions of the dosimeters. Response of the EPD-Ns to neutron radiation was evaluated using  $^{252}\text{Cf}$ , where measurements were made for thermal and fast neutrons using a polyethylene sphere as a moderator.

The gamma response was evaluated for two different energy levels using  $^{137}\text{Cs}$  and  $^{60}\text{Co}$ . X-ray photon response was evaluated at two different energies, 117 keV (beam code # H150), and 51 keV (beam code # M100).

## 2. Neutron Test

The response to thermal neutrons was tested using a  $^{252}\text{Cf}$  source moderated with a polyethylene sphere. The units were exposed to dose equivalent rates in the range of 10-120 mrem/hr. At 10 mrem/hr, response of the three units was very low with readings changing from 0-2 mrem/hr. At 50 mrem/hr the response remained low with readings changing between 0-4 mrem/hr. At 120 mrem/hr, all three units went into alarm. Meaningful readings could not be determined due to erratic indications.

**Conclusions:** All three units failed to measure neutrons, moderated and un-moderated, when exposed to  $^{252}\text{Cf}$  at various exposure rates. Follow up discussions with a technical representative of the manufacturer (Siemens) indicated that a site-specific calibration is required. According to the technical representative, all EPD-Ns are typically calibrated for the lowest energy limit, 0.025 eV. This raised some concerns about the validity of these devices to measure neutrons reliably when calibrated according to a specific site radiological characterization and then used at another site.

### 3. Gamma Test Using $^{137}\text{Cs}$

The units under test were exposed to gamma radiation from a  $^{137}\text{Cs}$  source. The source exposure rates and the response are listed in Table 1. Gamma-neutron cross talk measurements were also performed using the same source at one exposure rate only.

**Conclusions:** All three units appear to function satisfactorily in performing gamma measurements at various exposure rates. However, some gamma-neutron cross talk was detected when the units were switched to neutron mode while exposed to the  $^{137}\text{Cs}$  at one exposure rate. The results are shown in Table 2.

### 4. Gamma Test Using $^{60}\text{Co}$

A gamma test was also performed using  $^{60}\text{Co}$  at three different exposure/dose rates. Test results are shown in Table 3.

**Conclusions:** All units appear to function satisfactorily considering the accuracy of the calculated exposure rate of the source and the conversion factor used to determine the dose rate.

### 5. X-ray Test

Response to x-ray radiation was performed for two beam energies. One of the beams is code M-100 with 51 keV and the other is H-150 with 117 keV. The dose rates for the x-ray beams were determined using ion chamber measurements. Results are shown in Table 4.

**Conclusions:** All units appeared to perform satisfactorily.

**Table 1. Gamma measurement using  $^{137}\text{Cs}$**

| <b>Source Dose Rate<br/>(mrem/hr)</b> | <b>EPD-N #1<br/>(S/N 0635)<br/>Measured Dose<br/>Rate (mrem/hr)</b> | <b>EPD-N #2<br/>(S/N 0658)<br/>Measured Dose<br/>Rate (mrem/hr)</b> | <b>EPD-N #3<br/>(S/N 0683)<br/>Measured Dose<br/>Rate (mrem/hr)</b> |
|---------------------------------------|---|---|---|
| 303                                   | 300-340   | 300-350   | 290-320   |
| 503                                   | 500-550   | 470-590   | 470-580   |
| 1040                                  | 1000-1200   | 1000-1100   | 1000-1100   |
| 2505                                  | 2700-3000   | 2700-3000   | 2700-3000   |
| 8390                                  | 9300-9800   | 9500-10,000   | 9100-9500   |

**Table 2. Gamma-Neutron cross talk measurement ( $^{137}\text{Cs}$ )**

| <b>Source Dose Rate<br/>(mrem/hr)</b> | <b>EPD-N #1<br/>(S/N 0635)<br/>Measured Dose<br/>Rate (mrem/hr)</b> | <b>EPD-N #2<br/>(S/N 0658)<br/>Measured Dose<br/>Rate (mrem/hr)</b> | <b>EPD-N #3<br/>(S/N 0683)<br/>Measured Dose<br/>Rate (mrem/hr)</b> |
|---------------------------------------|---|---|---|
| 1040                                  | 0-15  | 11-22   | 13-25   |

**Table 3. Gamma test results for  $^{60}\text{Co}$**

| <b>Source Exposure<br/>Rate/Dose Rate<br/>(mrem/hr)</b> | <b>EPD-N #1<br/>(S/N 0635)<br/>Measured Dose<br/>Rate (mrem/hr)</b> | <b>EPD-N #2<br/>(S/N 0658)<br/>Measured Dose<br/>Rate (mrem/hr)</b> | <b>EPD-N #3<br/>(S/N 0683)<br/>Measured Dose<br/>Rate (mrem/hr)</b> |
|---|---|---|---|
| 75.8 / 77.9   | 97-100  | 92-100  | 100   |
| 561.9 / 577.24  | 600-700   | 660-740   | 630-740   |
| 2250 / 2311.43  | 2500-2700   | 2500-2700   | 2500-2700   |

**Table 4. X-ray measurement results**

|                            | <b>Calculated<br/>Dose Rate<br/>(mrem/hr)</b> | <b>EPD-N #1<br/>(S/N 0635)<br/>Measured<br/>Dose Rate<br/>(mrem/hr)</b> | <b>EPD-N #2<br/>(S/N 0658)<br/>Measured<br/>Dose Rate<br/>(mrem/hr)</b> | <b>EPD-N #3<br/>(S/N 0683)<br/>Measured<br/>Dose Rate<br/>(mrem/hr)</b> |
|----------------------------|---|---|---|---|
| <b>H-150 Beam<br/>Mode</b> | 513   | 490-550   | 490-550   | 460-550   |
|                            | 1999  | 1800-2000   | 1800-2000   | 1800-2000   |
|                            | 3610  | 2600-3000   | 2600-3000   | 2800-3000   |
| <b>M-100 Beam<br/>Mode</b> | 8280  | 7500-7600   | 7500-7600   | 7500  |



**APPENDIX A**

**EPD-N PUBLISHED SPECIFICATIONS**

## EPD-N PUBLISHED SPECIFICATIONS

- Real-time neutron and photon dosimetry
- Separate measurement and display of neutron and photon doses
- Sensitive to X and  $\gamma$ , and neutrons
- Multiple diode detectors
- Energy responses:
  - Photons: 50 keV – 7 MeV
  - Neutrons: 0.025 eV to > 10 MeV (thermal and intermediate)
- Lower limits of detection:
  - <1  $\mu$ Sv (100  $\mu$ rem)  $\gamma$ , X
  - ~ 10  $\mu$ Sv (1 mrem) n
- Dose rate response:
  - Neutron: 1 mrem/h to > 100 rem/h
  - Gamma: 0 mrem/h to 400 rem/h
- Estimation of neutron dose from high-level, short-duration exposure

**INTERNAL DISTRIBUTION**

1. P. J. Chiaro, Jr.
2. D. C. Gregory
3. J. B. Hunt
4. D. E. Perkins
5. P. D. Pruitt
6. K. L. Reaves
7. A. A. Shourbaji
8. R. F. Utrera
- 9 – 10. Central Research Library
11. ORNL Laboratory Records – RC

**EXTERNAL DISTRIBUTION**

12. Siraj Khan, U. S. Customs, Washington, DC
13. Randy Sullivan, Westinghouse Savannah River Company, Savannah River Site,  
Bldg. 735-B, Aiken, SC